

Schedule

ETS Solutions Asia Pte Ltd
1 Kaki Bukit Road 1
#03-20 Enterprise One
Singapore 415934

Certificate No. : LA-2017-0656-C
Issue No. : 3
Date : 30 November 2019
Page : 1 of 2

FIELD OF TESTING : Calibration and Measurement

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
<u>VIBRATION</u>		
Accelerometer		
1. <u>Voltage Sensitivity</u>		
1 Hz to < 4 Hz	ETS In-house Procedure: SOP-2017-CA-02 (Rev B, 19/09/2019)	4.0 %
4 Hz to 10 Hz		3.2 %
> 10 Hz to < 3000 Hz		2.0 %
3000 Hz to < 7000 Hz		3.2 %
7000 Hz to 10000 Hz		3.8 %
2. <u>Charge Accelerometer</u>		
1 Hz to < 4 Hz	ETS In-house Procedure: SOP-2017-CA-01 (Rev A, 01/03/2019)	4.6 %
4 Hz to 10 Hz		3.6 %
> 10 Hz to < 3000 Hz		2.3 %
3000 Hz to < 7000 Hz		3.5 %
7000 Hz to 10000 Hz		4.1 %
Electrodynamic Shaker Controller (AC Voltage , Charge)		
1. <u>Input Channel</u>		
1 Hz to < 10 Hz	ETS In-house Procedure: SOP-2017-CA-01 (Rev A, 01/03/2019)	0.4 %
10 Hz to < 2000 Hz		0.2 %
2 kHz to < 8 kHz		0.5 %
8 kHz to < 10 kHz		0.7 %
10 kHz to <15 kHz		1.2 %
15 kHz to 20 kHz		1.8 %
2. <u>Output Channel</u>		
1 Hz to < 10 Hz	ETS In-house Procedure: SOP-2017-CA-01 (Rev A, 01/03/2019)	0.4 %
10 Hz to < 10000 Hz		1.8 %
10 kHz to < 15 kHz		2.0 %
15 kHz to 20 kHz		2.0 %

Schedule



Certificate No. : LA-2017-0656-C

Issue No. : 3

Date : 30 November 2019

Page : 2 of 2

* CMC is expressed as an expanded uncertainty estimated at a level of confidence of approximately 95 %.

Approved signatories

Mr Miko Trisetya Judiutama

Note :

This laboratory is accredited in accordance with the recognised International Standard ISO/IEC 17025. A laboratory's fulfilment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and **management system requirements** that are necessary for it to consistently deliver technically valid calibration results. The **management system requirements** in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001.